



**INTERNATIONAL STANDARD ISO 11274:1998**  
**TECHNICAL CORRIGENDUM 1**

Published 2009-08-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Soil quality — Determination of the water-retention characteristic — Laboratory methods**

**TECHNICAL CORRIGENDUM 1**

*Qualité du sol — Détermination de la caractéristique de la rétention en eau — Méthodes de laboratoire*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 11274:1998 was prepared by Technical Committee ISO/TC 190, *Soil quality*.

*Page 2, 3.3*

The end of the 2nd line and part of the 3rd line should read : "..., e.g. 0 kPa to –20 kPa, –20 kPa to –100 kPa and –100 kPa to 1 500 kPa." instead of "..., e.g. 0 kPa to 20 kPa, 20 kPa to 100 kPa and 100 kPa to 1 500 kPa."

*Page 5, Table 2*

3rd column, the subheading should read "Surface of suction tables (–5 kPa matric pressure)".

4th column, the subheading should read: "Surface of suction tables (–11 kPa matric pressure)".

5th column, the subheading should read: "Surface of suction tables (–21 kPa matric pressure)".