
**Non-destructive testing — Radiation
methods for computed tomography —**

**Part 1:
Terminology**

*Essais non destructifs — Méthodes par rayonnements pour la
tomographie informatisée —*

Partie 1: Terminologie

STANDARDSISO.COM : Click to view the full PDF of ISO 15708-1:2017



STANDARDSISO.COM : Click to view the full PDF of ISO 15708-1:2017



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword.....	iv
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1

STANDARDSISO.COM : Click to view the full PDF of ISO 15708-1:2017

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html

This document was prepared by the European Committee for Standardization (CEN) (as EN 16016-1) and was adopted, under a special "fast-track procedure", by Technical Committee ISO/TC 135, *Non-destructive testing*, Subcommittee SC 5, *Radiographic testing*, in parallel with its approval by the ISO member bodies.

The first edition (ISO 15708-1:2002) having been cancelled and replaced by ISO 15708-2:2017, this second edition of ISO 15708-1 has been repurposed with a different title and scope and takes into consideration developments in computed tomography (CT) and computational power over the preceding decade.

A list of all parts in the ISO 15708 series can be found on the ISO website.

Non-destructive testing — Radiation methods for computed tomography —

Part 1: Terminology

1 Scope

This document gives the definitions of terms used in the field of computed tomography (CT). It presents a terminology that is not only CT-specific but which also includes other more generic terms and definitions spanning imaging and radiography. Some of the definitions represent discussion points aimed at refocusing their terms in the specific context of computed tomography.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

absorption

photoelectric absorption

mode of interaction between photons and matter whereby a photon is absorbed by an atom which then emits an electron whose kinetic energy is exactly equal to the energy-depleted photon's electron-binding energy

Note 1 to entry: See also *Compton scattering* (3.6).

3.2

angular increment

angular spacing between adjacent *CT projections* (3.12)

3.3

artefact

artificial feature which appears on the *CT image* (3.11) but does not correspond to a physical feature of the object

3.4

beam hardening

spectrum hardening

spectral change of a polychromatic beam caused by preferential attenuation of lower energy photons

Note 1 to entry: See also *cupping effect* (3.17).

3.5
calibration template
phantom

known reference object that is scanned to assess the performance of a *CT system* (3.15)

3.6
Compton scattering

mode of interaction between a photon and an electron, where the photon is scattered with reduced energy, and the difference of energy is transferred to the electron, also known as inelastic scattering or incoherent scattering

Note 1 to entry: See also *photoelectric absorption* (3.1).

3.7
computed tomography
CT

CT computed axial tomography
radiographic scanning technique that uses a number of *CT projections* (3.12) of an object at different angles in order to allow calculation of a *CT image* (3.11)

3.8
CT cone beam

scanning mode wherein each *CT projection* (3.12) is built from a set of *ray paths* (3.24) emanating from a point source and diverging in two dimensions, thereby forming a cone

3.9
CT data
CT dataset

CT projection (3.12) or *CT image* (3.11)

3.10
CT grey value
grey level

numerical value assigned to each *voxel* (3.30) in a *CT image* (3.11)

Note 1 to entry: This value represents the average *linear attenuation coefficient* (3.20) of the object volume for that voxel.

3.11
CT image
tomogram

2D or 3D image of the *CT grey values* (3.10) obtained by *reconstruction* (3.25)

3.12
CT projection

1D or 2D radiographic image

3.13
CT scan

set of relative movements between sample, source and detector, and the acquisition necessary to obtain a set of *CT projections* (3.12) that can be reconstructed into a *CT image* (3.11)

3.14
CT slice

2D *CT image* (3.11) with a finite thickness along a given plane

Note 1 to entry: See also *slice thickness* (3.29).

3.15**CT system tomograph**

equipment used to produce *CT images* (3.11)

3.16**CT volume**

3D *CT image* (3.11)

3.17**cupping effect**

feature due to *beam hardening* (3.4) in which the *CT grey values* (3.10) in a *CT image* (3.11) towards the centre of an homogeneous object are lower than those closer to the surface

3.18**density resolution**

measure of the extent to which a *CT image* (3.11) can be used to detect differences in the *linear attenuation coefficient* (3.20)

3.19**fan beam CT**

scanning mode wherein each *CT projection* (3.12) is built from a set of *ray paths* (3.24) emanating from a point source but considered to be diverging in only one dimension, thereby forming a 'fan'

3.20**linear attenuation coefficient**

X-ray attenuation (3.31) per unit path length of material at a given energy

Note 1 to entry: It is often expressed in cm^{-1} .

3.21**parallel beam CT**

scanning mode wherein each *CT projection* (3.12) is built from a set of parallel *ray paths* (3.24)

3.22**partial volume effect**

effect due to the finite voxel size of *CT images* (3.11) where properties of different materials are averaged within a single *voxel* (3.30)

3.23**pixel**

basic cell area in a 2D image or detector

Note 1 to entry: See also *voxel* (3.30).

3.24**ray path**

path that an X-ray travels from the source to a given *detector pixel* (3.23)

3.25**reconstruction**

process of transforming a set of *CT projections* (3.12) into a *CT image* (3.11)

3.26**region of interest****ROI**

sub-volume within an object or a *CT image* (3.11)

3.27

**region of interest CT
local tomography**

CT image (3.11) of a region of interest (ROI) (3.26) of an object using a set of CT projections (3.12) in which parts outside the ROI are not imaged in all of the CT projections (3.12)

3.28

sinogram

image formed by stacking vertically a set of 1D *CT projections (3.12)* from a complete set of angular positions in order of increasing projection angle

3.29

slice thickness

effective thickness of the X-ray beam in 2D tomography (i.e. that part of the X-ray beam that reaches the detector) measured at the centre of the object

3.30

voxel

volume element of a *CT image (3.11)* to which a *CT grey value (3.10)* is assigned

Note 1 to entry: It is the 3D equivalent of a *pixel (3.23)*.

3.31

X-ray attenuation

reduction in the intensity of X-rays as they pass through matter due to a combination of absorption and scattering

Note 1 to entry: See also *linear attenuation coefficient (3.20)*.

STANDARDSISO.COM : Click to view the full PDF of ISO 15708-1:2017